



Calibration certificate

Item No.: 500001/500002

Dimetix AG-Degersheimerstr. 14- CH-9100 Herisau - SWITZERLAND DIMETIX	
Certificate Of Calibration	
Certificate No.: 3200 Manufacturer: Dimetix AG Type: DLS-A 15 Description: DLS-A 15 Distance Laser Sensor Model No.: 500502 Serial No.: 44820102 Date of Calibration: 07-Januar-2005 Humidity: 4 @ ±0%RH Temperature: 20 ±0.4° C	
At the completion of the calibration, measured values were in specification stated in the technical reference manual.	
Calibration Equipment Used Model Number: D2002 Serial Number: 181144 Model Description: Laser Distance Sensor Calibration Date: 15.Januar.2005	
Herisau, 10-März-2005 _____ Andreas Zünd Quality Manager	

For the certificate of calibration of a DLS/FLS devices we gauge distances from 0.2 m to 16 m. For that, a interferometer ZLM 800 with an accuracy of 80nm is used as the reference measuring device. This device is traced to national and international standards.

Certificate of calibration – for new measuring devices DLS/FLS

By ordering new DLS sensors you may order at the same time to each DLS-device a certificate of calibration.

Recalibration with calibration certificate – for used measuring devices DLS/FLS

You may return already used DLS devices to Dimetix AG for recalibration. This units go through the whole quality process: Device cleaning, climatic exposure test cabinet, function test and calibration. Following to this the certificate of calibration will be created. Any repair are not included.

Order information

Item No.	Description	Packaging unit
500001	Calibration certificate for new measuring devices DLS/FLS	1 piece
500002	Recalibration with calibration certificate for used measuring devices DLS/FLS	1 piece

12. Nov. 2009 / Subject to change without notice

500001_002 Calibration certificate_EN_V2.odt